

CORRIGENDA

Technical specifications of personal protective equipment for COVID-19: interim guidance, 13 November 2020

(WHO/2019-nCoV/PPE_specifications/2020.1)

Page 3, Table (1), Column 2, lines 7–10

Delete:

Quality management system from the manufacturer, for the PPE types	Certified quality management system for medical devices (e.g. ISO 13485) and application of risk management to medical devices (e.g. ISO 14971). General quality management (e.g. ISO 9001) (for non-medical devices). European Union (EU) Module C2 or D conformity to type certificate (Category III CE certified PPE only).
---	--

Insert:

Quality management system from the manufacturer, for the PPE types	Certified quality management system for medical devices (e.g. ISO 13485) and application of risk management to medical devices (e.g. ISO 14971), if applicable. General quality management (e.g. ISO 9001) (for non-medical devices). European Union (EU) Module C2 or D conformity to type certificate (Category III CE certified PPE only).
---	---

* Page 4, Table (1), Column 3, lines 3–6 *** NEW 2020-12-21 ***

Delete:

EN 455 EN 374, optional additional: ASTM D6319, D3578, D5250, D6977 Or alternative equivalent set of standards

Insert:

EN 455 ASTM D6319, D3578, D5250, or D6977 EN 374, optional additional Or alternative equivalent set of standards

Page 9, Table (2) ‘Medical face masks’, rows 4–6

Delete:

Synthetic blood penetration (kPa)	120 mmHg (ISO 22609)	80 mmHg or 10.7 kPa (Level 1)	120 mmHg or 16 kPa	N/A
Microbial cleanliness (CFU/g)	16 kPa (Type IIR)	120 mmHg or 16 kPa (Level 2) 160 mmHg or 21.4 kPa (Level 3)		
Total inward leakage	≤ 30	N/A	≤ 100	≤ 100

Insert:

Synthetic blood penetration (kPa)	120 mmHg (ISO 22609) 16 kPa (Type IIR)	80 mmHg or 10.7 kPa (Level 1) 120 mmHg or 16 kPa (Level 2) 160 mmHg or 21.4 kPa (Level 3)	120 mmHg or 16 kPa	N/A
Microbial cleanliness (CFU/g)	≤ 30	N/A	≤ 100	≤ 100

These corrections have been incorporated into the electronic file.